

A WEARABLE ELECTRONIC MONITORING DEVICE FOR LOW PRESSURE GARMENT
APPLICATIONS AND TEMPERATURE ANALYSIS FOR PREVENTION OF
ULCERATION AND INFECTION

By

Francisco Mata Carlos

Bachelor of Science in Electrical Engineering
University of Nevada, Las Vegas
2020

A thesis submitted in partial fulfillment
of the requirements for the

Master of Science in Engineering – Electrical Engineering

Department of Electrical and Computer Engineering
Howard R. Hughes College of Engineering
The Graduate College

University of Nevada, Las Vegas
May 2022

© Copyright 2022 by Francisco Mata Carlos

All Rights Reserved



Thesis Approval

The Graduate College
The University of Nevada, Las Vegas

April 8, 2022

This thesis prepared by

Francisco Mata Carlos

entitled

A Wearable Electronic Monitoring Device for Low Pressure Garment Applications and
Temperature Analysis for Prevention of Ulceration and Infection

is approved in partial fulfillment of the requirements for the degree of

Master of Science in Engineering – Electrical Engineering
Department of Electrical and Computer Engineering

R. Jacob Baker, Ph.D.
Examination Committee Chair

Kathryn Hausbeck Korgan, Ph.D.
*Vice Provost for Graduate Education &
Dean of the Graduate College*

Grzegorz Chmaj, Ph.D.
Examination Committee Member

Yahia Baghzouz, Ph.D.
Examination Committee Member

Ashkan Salamat, Ph.D.
Graduate College Faculty Representative

ABSTRACT

The healing process of several medical conditions related to venous insufficiency and burn scar management can be improved by the proper application of pressure. Current practices for applying pressure to an extremity do not guarantee the consensus ideal interface pressures for such medical conditions. Applying a low pressure range (within 15 – 40 mm Hg) is challenging due to many factors, but especially because the actual pressure applied by a compression garment used on a patient is unknown. The proposed device, used to sense and display low pressure values, is described in detail throughout the chapters of this thesis.

An electronic device was designed, fabricated, assembled, calibrated, and tested in order to deliver a small wearable electronic system which can measure and provide low pressure values inside the range of interest. The sensing unit is intended to be placed underneath the compression garment against the arm or leg of the subject and take force (which is then converted to pressure) and temperature readings. These readings are analyzed by the microcontroller and sent via Bluetooth to the smart phone application where they can be reviewed by the user to make appropriate adjustments to the applied pressure, if necessary. The purpose of the thermal sensor is to sample the skin temperature over a period of time and use the collected data to determine the potential presence of skin infections. The four major elements that complete this system are the sensor unit, the controlling unit, the Bluetooth module, and the smart phone application, which can be powered by a single 3.7V rechargeable battery lasting tens of hours without replacement. The system can be adapted to run for several days if timers are implemented. The test results conducted to the wearable gadget show exceptional linearity and accuracy within 1N range (75 mm Hg), providing a promising solution for sensing low pressure values.

ACKNOWLEDGEMENTS

First, I would like to thank my advisor (and chair of my advisory committee), Dr. Baker, for the immense support and guidance he has provided me, which has led to my professional and academic success. The skills that I have obtained and opportunities that have become available to me have all been the result of being part of his research team. The path that I have taken in my engineering career has been remarkably influenced by Dr. Baker since the first time I took an electrical engineering course. Dr. Baker was the instructor for my first circuits class where his teaching method sparked my fascination with circuits and electronics. Since that moment, my academic performance and proficiency have been constantly improving, leading to this stage of my life.

I would also like to thank the other members of my advisory committee – Dr. Baghzouz, Dr. Greg, and Dr. Salamat – for all the knowledge they have provided me through their courses I have taken over the years. The education I received from Dr. Baghzouz and Dr. Greg have prepared me to work on a set of new electronic devices that are related to the power grid, and also require IoT technology. Dr. Salamat on the other hand has always inspired me to dig deep to really understand the mathematical perspective of the physical world, initially making physics my favorite subject, before I met Dr. Baker. Nevertheless, it was due to Dr. Salamat's educational influence that I decided to stay in the science and engineering career path.

I would like to thank James Skelly, my colleague and a very special friend, who I have always looked up to for having many outstanding qualities, which relate to academic, social, and physical activities. Since I've met him, he has managed to impress me with the results from every obstacle he has encountered. He is the kind of person that dives deep into a subject to define a problem from a mathematical and physical perspective. I am privileged to have worked alongside

him throughout several projects within the Baker research group. I would also like to thank all the previous and current research students who have impacted my life, including Dr. Sachin Namboodiri, Daniel Senda, Mario Valles, David Santiago, Chris Barr, Jazmine Bolor, Abraham Lopez and Abraham Castaneda.

A special thanks to Dr. Todd Meyrath for his continuous professional and academic advice. Dr. Meyrath's guidance has encouraged me to push further in my academic career, which is another reason I decided to pursue a master's degree in electrical engineering. He is an exceptional person and I am honored to be working for his company. Another special thanks to my whole family for their support and especially my parents, for always showing me through their actions that hard work pays off.

TABLE OF CONTENTS

ABSTRACT.....	iii
ACKNOWLEDGEMENTS.....	iv
TABLE OF CONTENTS.....	vi
LIST OF TABLES.....	viii
LIST OF FIGURES	ix
CHAPTER 1: INTRODUCTION	1
1.1 BACKGROUND.....	1
1.2 PROBLEMS WITH CURRENT COMPRESSION METHODS	5
1.3 PROPOSED METHOD AND OBJECTIVES	7
CHAPTER 2: DESIGN OF ELECTRONIC MONITORING DEVICE	9
2.1 DESIGN OVERVIEW	9
2.1.1 PRINCIPLE OF OPERATION	10
2.2 THE SENSOR UNIT	10
2.2.1 FORCE SENSOR.....	11
2.2.2 TEMPERATURE SENSOR.....	14
2.2.3 PCB DESIGN OF SENSOR UNIT	15
2.2.4 SPRING-LOADED PINS (POGO PINS)	17
2.2.5 SENSOR UNIT ASSEMBLY	18
2.3 CONTROLLING UNIT.....	25
2.3.1 PIC18 MICROCONTROLLER	26
2.3.2 THE INSTRUMENTATION AMPLIFIER	27
2.3.3 BUCK-BOOST CONVERTER.....	28

2.4 BLUETOOTH LOW ENERGY (BLE) MODULE	30
2.4.1 BLUETOOTH PRINCIPLE OF OPERATION.....	30
2.4.2 BLUETOOTH HM-10 MODULE	31
2.5 DESIGN OF SMART PHONE APPLICATION.....	33
2.5.1 WEB-BASED APPLICATION IDE	33
2.5.2 COMPOSITION OF MOBILE PHONE APPLICATION STRUCTURE.....	34
CHAPTER 3: FLOW OF DATA AND COMMUNICATION	38
3.1 MPLAB X IDE SOFTWARE AND PROGRAMMER.....	38
3.2 MPLAB C CODE	39
3.2.1 THE SENSING STAGE.....	40
3.2.2 DATA PROCESSING AND DATA LABELING STAGE	43
3.2.3 DATA ARRANGEMENT STAGE.....	50
3.3 MIT APP INVENTOR CODE.....	52
CHAPTER 4: PROTOTYPE AND PERFORMANCE RESULTS.....	58
4.1 PROTOTYPE AND DEVICE COST BREAKDOWN	58
4.2 PROTOYPE PERFORMANCE TEST RESULTS.....	61
CHAPTER 5: CONCLUSION AND FUTURE WORK.....	67
5.1 CONCLUSION	67
5.2 FUTURE WORK	68
APPENDIX I: COMPLETE MPLAB X C CODE	73
APPENDIX II: COMPLETE MIT APP INVENTOR BLOCK CODE	78
REFERENCES	80
CURRICULUM VITAE.....	84

LIST OF TABLES

Table 2.1	Power modes for CC2541 Bluetooth SoC.....	33
Table 2.2	Phone application functionality list	37
Table 4.1	Cost breakdown of sensor unit	60
Table 4.2	Cost breakdown of controlling unit with Bluetooth included	60

LIST OF FIGURES

Figure 1.1	Stages of chronic venous insufficiency (CVI) [10]	3
Figure 1.2	Keloid scar with finite element analysis (FEA) showing tension levels [16]	5
Figure 1.3	Patient pressure value reading over the course of several months [11]	6
Figure 1.4	Pressure ulcer due to improper stocking size [18]	7
Figure 2.1	System design concept and flow of communication.....	9
Figure 2.2	System sensor unit (Solidworks model)	11
Figure 2.3	a) Force sensor and b) its circuit diagram [20]	13
Figure 2.4	Wheatstone bridge circuit	13
Figure 2.5	Temperature sensor size and package [21]	15
Figure 2.6	Current and output voltage vs. ambient temperature [21]	15
Figure 2.7	Sensor unit top and bottom PCBs with schematics	16
Figure 2.8	Top and bottom 3D PCB models of sensor unit	17
Figure 2.9	Spring-loaded pin [22]	18
Figure 2.10	Exploded view sensor unit assembly components.....	19
Figure 2.11	Sensor unit pressing and depressing operation	20
Figure 2.12	Plot of output voltage vs. mass	21
Figure 2.13	Plot of output voltage vs. force	21
Figure 2.14	Plot of output voltage vs. force with initial force removed	22
Figure 2.15	Temperature calibration setup of sensor unit.....	23
Figure 2.16	Plot of temperature chamber vs. temperature unit output voltage	24
Figure 2.17	Plot of IR thermometer and sensor unit temperature test	24
Figure 2.18	Temperature comparison of IR thermometer vs. unit sensor after optimization.....	25

Figure 2.19	Main controlling unit schematic with Bluetooth included.....	26
Figure 2.20	Buck-boost efficiency plot.....	29
Figure 2.21	HM-10 Bluetooth module [27]	31
Figure 2.22	BLE pin configurator and typical connections [27].....	32
Figure 2.23	Designer view window of MIT App Inventor	34
Figure 2.24	Smart phone application interacting with the system	36
Figure 3.1	MPLAB C code flowchart	40
Figure 3.2	Flowchart sensing stage portion.....	42
Figure 3.3	ADC conversions of force, temperature, and battery charge readings	43
Figure 3.4	Data processing and labeling stage	44
Figure 3.5	Obtaining voltage, force, and pressure values from their respective functions	45
Figure 3.6	Function to calculate the force from ADC readings	45
Figure 3.7	Function to calculate pressure (mm Hg) from the force values.....	46
Figure 3.8	Obtaining voltage and temperature values from their respective functions	47
Figure 3.9	Function that resets the ADC to take new analog input samples.....	47
Figure 3.10	Function that computes the temperature from the sensed signal	48
Figure 3.11	Computing battery voltage values from ADC readings.....	49
Figure 3.12	Formulation of data labels to be use by smart phone app.....	50
Figure 3.13	Organization of analyzed data prior to Bluetooth transmission.....	51
Figure 3.14	The combination of the data values and their assigned labels	51
Figure 3.15	Transferring data to the Bluetooth module	52
Figure 3.16	MIT App Inventor software windows: Blocks and Designer	53
Figure 3.17	Smart phone app flowchart	54

Figure 3.18	Removing unused characters from receiving data	55
Figure 3.19	Management of received data for proper label classification	56
Figure 3.20	Low voltage battery condition statement.....	56
Figure 3.21	Management of received data for appropriate storage options	57
Figure 4.1	Complete wearable device system	59
Figure 4.2	Force comparison test: wearable device vs. actual force with offset.....	63
Figure 4.3	Force comparison test: wearable device vs. actual force with no offset.....	63
Figure 4.4	Pressure comparison test: wearable device vs. calculated pressure with offset	64
Figure 4.5	Pressure comparison test: wearable device vs. calculated pressure with no offset	64
Figure 4.6	Temperature comparison test (1 st subject): wearable device vs. IR thermometer ..	65
Figure 4.7	Temperature comparison test (2 nd subject): wearable device vs. IR thermometer ..	65
Figure 4.8	Battery life expectancy with one sensor	66
Figure 4.9	Battery life expectancy simulation with 10 sensors.....	66
Figure 5.1	Improved version of sensor unit	69
Figure 5.2	Low profile PCB connector	71
Figure 5.3	Comparison of BLE module sizes, HM-10 (left), HM-10 (middle), BM71 (right)	72
Figure A2.1	Smart phone app setup code	78
Figure A2.2	Smart phone data management code.....	79

REFERENCES

- [1] R. Weiss et al, “Venous insufficiency,” USENET: drug and diseases.dermatology, Sep. 25, 2020 [Accessed Feb. 8, 2022].
- [2] M. Porter, “A Case Study of Venous Leg Ulceration,” in *British Journal of Community Nursing*, Aug. 2018. [Online]. Available: <https://www.magonlinelibrary.com/doi/full/10.12968/bjcn.2018.23.Sup9.S30>.
- [3] I. C. Valencia et al, “Chronic venous insufficiency and venous leg ulceration,” *Journal of the American Academy of Dermatology*, vol. 44, no. 3, pp. 401-404, Mar. 2001. Available: ScienceDirect, <https://www.sciencedirect.com/science/article/pii/S0190962201031553#aep-section-id37>.
- [4] C. V. Ruckley, “Socioeconomic impact of chronic venous insufficiency and leg ulcers.” *Sage Journals*, vol. 48, no. 1, pp. 67-69, 1997.
- [5] “Chronic Venous Insufficiency,” The John Hopkins University, Baltimore, Maryland, United States. Accessed: Feb. 8, 2022. [Online]. Available: <https://www.hopkinsmedicine.org/health/conditions-and-diseases/chronic-venous-insufficiency>.
- [6] B. Eklof et al, “Revision of the CEAP classification for chronic venous disorders: consensus statement,” *Journal of Vascular Surgery*, vol. 40, no. 6, pp. 1248-1252, Dec. 2004. Available: ScienceDirect, <https://www.sciencedirect.com/science/article/pii/S0741521404012777>.
- [7] R. Weiss et al, “Venous insufficiency treatment & management,” USENET: drug and disease.dermatology, Sep. 25, 2020. [Accessed Feb. 8, 2022].

- [8] J. Menezes, MD. (2021). Optimization of Compression Therapy - Smart Stockings
[PowerPoint slides].
- [9] J. C. Mayberry et al, "Fifteen-year results of ambulatory compression therapy for chronic venous ulcers," Oregon Health Sciences University, Portland, Oregon, United States, PMID: 2020902, May 1, 1991. Accessed: Feb. 9, 2022. [Online]. Available: <https://europepmc.org/article/med/2020902>.
- [10] E. Conde, "CEAP Classification of Chronic Venous Disorders: Let's all speak about the same language," *elenaconde.com*, Jan. 20, 2019. [Online]. Available: <https://www.elenaconde.com/en/ceap-classification-of-chronic-venous-disorders-lets-all-speak-the-same-language/>. [Accessed Feb. 14, 2022].
- [11] J. Wiseman et al, "Variability of pressure at the pressure garment-scar interface in children after burn: A pilot longitudinal cohort study," *Burns*, vol. 45, no 1, pp. 103-113, Feb. 2019. Available: ScienceDirect, <https://www.sciencedirect.com/science/article/pii/S0305417918302596>.
- [12] Wound Source. (2022). *Scar Management* [Online]. Available: <https://www.woundsource.com/patientcondition/scar-management>
- [13] American Academy of Dermatology Association. (2022). *Scars: Signs and Symptoms* [Online]. <https://www.aad.org/public/diseases/a-z/scars-symptoms>
- [14] C. Roques, "Pressure therapy to treat burn scars," *Wound Repair and Regeneration*, vol. 10, pp. 122-125, May 2002.
- [15] S. Chris, "The true cost of burn," *Burns*, vol. 38, no 7, pp. 967-974, Nov. 2012. Available: ScienceDirect <https://www.sciencedirect.com/science/article/pii/S030541791200174X#tbl0005>.

- [16] O. Rei, A. Satoshi, “Endothelial dysfunction may play a key role in keloid and hypertrophic scar pathogenesis – Keloids and hypertrophic scars may be vascular disorders,” *Medical Hypotheses*, vol. 96, pp. 51-60, Nov. 2016. Available: Elsevier <https://www.sciencedirect.com/science/article/pii/S0306987716302766#f0020>.
- [17] S. Rathbun, A.Kirkpatrick, “Treatment of chronic venous insufficiency,” *Current Treatments Options in Cardiovascular Medicine*, vol. 9, pp 115-126, May 2007. Available: SpringerLink <https://link.springer.com/article/10.1007/s11936-007-0005-6>.
- [18] J. Black, P. Alves, C. Brindle, “Use of wound dressings to enhance prevention of pressure ulcers caused by medical devices,” *International Wound Journal*, vol. 12, no. 3, pp. 322-327, Jul. 2013. Available: Wiley Online Libraries <https://onlinelibrary.wiley.com/doi/full/10.1111/iwj.12111>.
- [19] M. Fierheller and R. Sibbald, “A Clinical Investigation into the Relationship between Increased Periwound Skin Temperature and Local Wound Infection in Patients with Chronic Leg Ulcers,” *Advances in Skin & Wound Care*, vol. 23, no. 8, pp. 369-379, Aug. 2010. Available: Ovid <https://oce.ovid.com/article/00129334-201008000-00008/HTML>.
- [20] Alps Alpine, “Force Sensor,” HSFPAR003A datasheet, Jan. 2018 [Revised Mar. 2019].
- [21] Maxim, “High-Slop, Low-Power, Analog Temperature Sensor in an SC70 Package,” MAX6612MXK datasheet, Jul. 2002.
- [22] Amazon. (1996-2022). *Mxfans 100xGold-Plated 6mm Copper Probes Spring pogo Pin Connector* [Online]. Available: <https://www.amazon.com/Mxfans-Gold-plated-Copper-Needles-Connector/dp/B07FQF2BRC>.
- [23] TestEquity, “Temperature Chamber,” Model 105 datasheet, Feb. 2002.

- [24] Microchip, “High-Performance Microcontroller with XLP Technology,” PIC18(L)F26K22 datasheet, 2010-2012.
- [25] Maxim Integrated, “Ultra-low Offset/Drift, Precision Instrumentation Amplifiers with Buffer,” MAX4208 datasheet, May 2015.
- [26] Linear Technology, “200mA Buck-Boost Synchronous DC/DC Converters,” LT3531 datasheet, Aug. 2007.
- [27] DSD TECH, “Bluetooth module,” HM-10 datasheet, Jan. 2001.
- [28] MILL-MAX, “SPRING-LOADED PIN,” 0965 datasheet.
- [29] Microchip, “16-Bit, Multi-Channel delta-sigma ADC with I2C Interface and On-Board Reference,” MCP3426 datasheet, 2009.
- [30] DIODES, “300mA RF ULDO REGULATOR,” AP2210 datasheet, Mar. 2021.
- [31] Hirose Electric Co.,LTD, “Low Profile Connector,” DF58 Series datasheet, Nov. 2016.
- [32] J. Skelly, “Monitored Compression Therapy: Using Smart Technology to Optimize the Treatment of Lower Extremity Swelling,” M.S. thesis, Dept. of ECE, UNLV, Las Vegas, NV, 2022.

CURRICULUM VITAE

Graduate College

University of Nevada, Las Vegas

Francisco Mata Carlos

Contact Information:

Email: fjmata05@gmail.com

Degree:

Bachelor of Science, Electrical and Computer Engineering, 2020

University of Nevada, Las Vegas

Employment:

Vorpall Research Systems, LLC 2019 – Present

University of Nevada, Las Vegas (UNLV) 2017 – Present

Thesis Title:

A Wearable Electronic Monitoring Device for Low Pressure Garment Applications and
Temperature Analysis for Prevention of Ulceration and Infection